

# CXDI-80C Wireless\*

Digital Radiography System



## *Make Way for the World's First Compact Wireless Flat Panel Detector*

Feel the freedom that comes with *WIRELESS* Digital Radiography. The CXDI-80C Wireless DR Detector is the first compact wireless flat panel detector delivering high resolution images at 125 Micron Pixel pitch.

## *Designed to Work In Smaller Places*

Only 5.3\*\* lbs. and an imaging area of 11" x 14"\*\*, the CXDI-80C Wireless DR Detector is durable, super compact and lightweight. Coupled with a long battery life and seamless data transfer from detector to console, the CXDI-80C Wireless DR Detector is the ideal size for fitting into most Neonatal Intensive Care Units (NICU) beds without modification.

## *Canon Technology Throughout*

The CXDI-80C Wireless DR Detector incorporates Canon made LANMIT® (Large Area New Metal-Insulator Semiconductor) technology and a proprietary glass substrate for higher resolution, high contrast images. In addition, the CXDI-80C Wireless DR Detector includes a CsI scintillator that optimizes light channeling and produces high quality images at a lower X-ray exposure to the patient.

Go from room-to-room freely without cables and experience the higher sensitivity, lower dose and dependability of the Canon CXDI-80C Wireless Compact DR Detector.

**Canon**

# CXDI-80C Wireless\*

## Digital Radiography System



### Specifications

<b>IMAGE SIZE</b> <b>11 x 14**</b> INCHES	<b>WEIGHT</b> <b>5.3**</b> POUNDS	<b>PIXEL PITCH</b> <b>125</b> MICRONS	<b>IMAGE PREVIEW</b> <b>3</b> SECONDS
---	---	---	---

Scintillator	CsI (CsI:TI)
Purpose	General radiography
Method	Cassette size detector, scintillator & amorphous silicon (a-Si)
Sensor	LANMIT (Large Area New-MIS sensor and TFT)
Pixels	2,192 x 2,800 pixels (6.1 Megapixels)
A/D	14-bit
Grayscale	4,096 grayscale (12-bit)
Wireless Standard	IEEE 802.11N
High Resolution Image Display Time	Approx. 5 seconds after X-ray exposure
Cycle Time	Approx. 15 seconds after X-ray exposure
DICOM®	DICOM 3.0 compatible, Print Management Service Class (SCU), Storage Service Class (SCU), and others
Battery Performance	Up to 140 images
Recharging Time	Less than 3 hours
Operating Environment	Sensor unit: 41-95°F (5-35°C) Humidity 30 - 80% RH (non-condensing)
Dimensions (L x W x D)	12.1" x 15.1" x 0.6" in (307 x 384 x 15mm)

\* FDA 510(k) Pending

\*\* Imaging Size and Weight are approximate figures.

*Accelerating inSight.*

08/2011 AHRA

CANON and CXDI are registered trademarks of Canon Inc. in the United States and may also be registered trademarks or trademarks of other countries. DICOM is a registered trademark of the National Electrical Manufacturer's Association for its standards publications relating to digital communications of medical information. All other referenced marks are trademarks of their respective owners and are hereby acknowledged. Specifications are subject to change without notice.

© 2011 Canon U.S.A., Inc

CANON MEDICAL SYSTEMS  
A Division of Canon U.S.A., Inc.

15955 Alton Parkway  
Irvine, CA 92618-3616, U.S.A

Telephone: (800)970-7227, (949)753-4160

Fax: (949)753-4164

Email: drsales@cusa.canon.com

Website: www.usa.canon.com/cxdi-80cwireless